

# DIGITAL MULTIMETER KEW 1051 / 1052 / 1061 / 1062

# The Best Reliable Multimeters with Safety Terminal Shutters



KYORITSU ELECTRICAL INSTRUMENTS WORKS, LTD. www.kew-ltd.co.jp

# **High Accuracy, High Performance and Reliable Measurements**

#### Top class Accuracy

- $\cdot\,0.02\%$  basic DC accuracy for 1061/1062.
- $\cdot\,0.09\%$  basic DC accuracy for 1051/1052.

#### Dual Display

- 1061/1062 : 50000 counts, 51 segments bar graph with white backlight display.
- 1051/1052 : 6000 counts, 31 segments bar graph with white backlight display.

#### Wide AC Frequency Bandwidth #1061, 1062 only

- · 1062 : ACV frequency bandwidth from 10Hz to 100kHz.
- · 1061 : ACV frequency bandwidth from 10Hz to 20kHz.

# **Advanced Functions**

#### User calibration function

- $\cdot$  Calibration and adjustment are possible by simple operation of DMM keys.
- New technology enables the adjustment for the frequency bandwidth characteristic. #1061, 1062 only #A calibrator is necessary for calibration.

#### Low-pass Filter \*except for 1061

- AC measurement can be limited to low frequency, helping for instance voltage measurements in the presence of variable speed motor drivers or inverters.
- · The Low-pass filter can be switched ON/OFF.

#### LowPower-Ω measurement \*1062 only

 This function uses a test voltage which is lower than 0.7V (that is the typical junction voltage drop of semiconductors) thus it allows testing of resistors on a circuit board without unsoldering them.

#### Selection of the reading mode \*1052, 1062 only

• Selectable True RMS or MEAN measurement. The presence of distortion in an AC signal can be confirmed, if the measured True RMS and MEAN values are different.

#### Sensor mode \*1051, 1052 only

• The DMM measures the output voltage of an external sensor (e.g. clamp sensor, light sensor, temperature sensor, etc.) in the secondary display, while the primary display can be set to show the unit of the measured parameter (e.g. A, mA, Lux, °C) according to the conversion ratio chosen.

#### Peak Hold function \*1062 only

- · Response time : 250µs
- The instantaneous peak values can be easily captured where normally it is impossible by MIN/MAX/AVG function.

#### Auto Hold function

• The measured value is held on the display just by removing the test leads from the circuit under test. Users can remain safely concentrated on the measuring point without the need to press the Hold key.

#### Relative and Percentage calculation

• Can calculate and display Relative values or Percentage (%) against the reference measurement values.

#### True RMS Measurement


 Ensures accurate readings, avoiding errors (of up to 50%) which can occur when non-sinusoidal waveforms, created by common non linear loads such PCs, Inverters, switch-mode power supplies, etc, are measured.

#### DC+AC True RMS Measurement \*1061, 1062 only

- Accurate AC True RMS measurements also in the presence of superimposed DC component.
- · AC and DC values are displayed simultaneously via dual display.







#### Minimum / Maximum / Average function \*\* except for 1051

- Can record the MIN/MAX/AVG values during the measurement process displaying the data and the elapsed time.
- %The average value is shown by dividing the integrated record data by the number of recording time.
- Duty cycle ratio measurement \*1061, 1062 only
  - The duty cycle ratio is displayed in percentage (%).

#### Decibel dBV, dBm calculation \*1061, 1062 only

Can perform logarithmic calculations on AC voltage.
 «Reference resistance value:
 4/8/16/32/50/75/93/110/125/135/150/200/250/300/500/600/800/900/1000/1200Ω

# Safe and Durable Design. Wide Operating Temperature.

# Complies with IEC 61010-1 CAT IV 600V / CAT II 1000V Safety shutters to prevent incorrect test leads insertion in current terminals

•Terminal shutters are opening or closing being linked with the rotation of the function switch.

#### **Operation of the Safety Shutters**

Safety shutters are open or closed when the appropriate function is selected because they are linked with the rotation of the function switch.





If the DMM has the function switch in position 1 (V,  $\Omega$ , TEMP, etc) the safety shutters close the input terminals for the current measurements (µA, mA, A) and then the test leads cannot be plugged-in.

If the DMM has the function switch in position 2 (current measurements) then the safety shutters automatically open making it possible to plug-in the test leads in the input terminals for the current measurements (µA, mA, A).

#### Very wide operating temperature range

- · From -20 to +55℃ for 1061/1062
- From -10 to +55°C for 1051/1052

#### High specs UL standard fuses for extra safety

· Fuses rated at 1000V with 30kA of breaking capacity.

#### Over molding case

• Made by "Elastomer", a superior shock sustainable material. Perfectly fits to hand.

# Comprehensive support for data management \*\*\*\*cept for 1051

#### Large internal memory to store test data

- · 1062: 10,000 data in Logging mode, 100 data manually saved.
- · 1061: 1,000 data in Logging mode, 100 data manually saved.
- · 1052: 1,600 data in Logging mode, 100 data manually saved.
- · Logging interval can set from 1 sec. to 30 min.

#### Test data can be transferred to a PC or directly to a Printer\*


- · Real-time data can be transferred and shown on a PC.
- Real-time transferring permits the saving of a considerable amount of data on a PC.

·Stored data of internal memory can be monitored by PC.

#### Data management with the software DMM Application\*

- · List of measured data can be converted into Graph.
- $\cdot$  Data can be transferred to Excel\*\* and saved as CSV file.
- \*Optional accessories are required, refer to last page.

\*\*Excel is a registered trademark of Microsoft in the US.




To protect us against overvoltage spikes, we must use instruments that meet the requirements for high protection standards.

The IEC (International Electrotechnical Commission) has prepared an International and European safety standard named IEC 61010-1 with the aim of defining the safety requirements for measuring instruments.

In particular IEC 61010-1 standard defines also the safety Measurement areas called Categories, shortly indicated with the abbreviation "CAT".

These Categories start from O to CAT IV and the most dangerous one is the CAT IV. The figure above shows some area examples of Measurement Categories.

Measurement category	Description	Examples
0	Measuring circuits without a MEASUREMENT CATEGORY.	Signal level circuits of electronic PCBs, etc.
CAT II	For measurements performed on circuits directly connected to the low voltage installation.	Appliances, portable equipment, ect.
CAT II	For measurements performed in the building installation.	Distribution board, circuit breaker, ect.
CAT Ⅳ	For measurements performed all the source of the low-voltage installation.	Overhead wire, cable systems, ect.



#### Versatile Digital Multimeters KEW 1051/1052

#### General Specification

Measurement function:	DC Voltage, AC Voltage, DC Current, AC Current, Resistance, Frequency,	Measurement cycle:	5 times per second (except frequency measurement : one time per second, Resistance
	Temperature, Capacitor, Continuity Check, Diode Test		measurement (6/60M $\Omega$ ) : 2.5 times per second, capacitor measurement (1000µF)
	Effective value (root mean square value) detection (True RMS) and mean value detection		max.0.14 time per second)
	(MEAN) can be switched during AC voltage measurement (1052 only).		Bar graph display approx. 25 times per second (at AC, Ω)
	The low-pass filter can be switched on/off during AC voltage or AC current measurement.	Operating temperature	and humidity range:
Other functions: Data Hole	d (D·H), Auto Hold (A·H), Range Hold (R·H), Maximum value* (MAX),		-10 to 55℃, 80%RH or less (no condensation) 70%RH or less at 40 to 55℃.
	Minimum value* (MIN), Average value* (AVG), Zero Adjustment (Capacitor, Resistance),	Storage temperature ar	nd humidity range: -30 to 70°C, 70%RH or less (no condensation)
	Relative values, Save to Memory*, Auto Power Off (Approx. 20 minutes), LCD backlight. *: For	Temperature coefficient	t: (Accuracy at 23±5℃×0.1)/℃ should be added.
	model KEW1052 only		(Temperature range: −10 to 18°C and 28 to 55°C)
Display:	4-digit (LCD)7-segment	Power supply:	R6/LR6(1.5V): 4
	Main-display6000 counts	Battery life:	Approximately 300 hours (Operating hours of alkaline batteries when in DC voltage-mode.)
	Sub-display		Note: The battery life varies depending on the operating conditions.
	Bar graph indicator31-segment	Withstand voltage:	6.88kV rms AC for five seconds (across input terminals and casing)
	Polarity Indicator:"—" Appears automatically when the polarity is negative.	Dimension:	Approx. 192(L)×90(W)×49(D)mm
	Overrange Indicator " OL "	Weight:	Approx. 570g (including batteries and blank cover)
	Low-battery Indicator" 🚛 " Appears when the batteries become low.	Applicable standards: Accessories:	IEC 61010-1 CAT IV 600 <sup>7</sup> / CAT III 1000V Pollution degree 2, IEC 61010-031, IEC 61326-1 Batteries : 4, Test leads: 1set (7220A), Fuse (included): 440mA/1000V (8926), 10A/1000V (8927), Blank cover, Instruction manual: 1

#### Specification

Test conditions: Temperature and humidity: 23±5°C at 80%RH or less Accuracy: ± (% of reading + digits) Note: Each response time is a value to rated accuracy within selected range.

DC Voltage Measurement(....V)

Range	Accuracy	Input Impedance	Overload Protection
600.0mV		10ΜΩ	
6.000V	0.09+2	11MΩ	1000V DC
60.00V	0.09+2		1000V DC
600.0V		10ΜΩ	1000v rms AC
1000V	0.15+2		

NMRR: 60dB or more 50/60Hz ± 0.1% CMRR: 120dB or more 50/60Hz (Rs=1kΩ)

Response time: 1 sec max.

#### AC Voltage Measurement (~V) AC Coupling: True RMS value delection, sine wave MEAN value delection and True RMS value calibration (KEW1052 only)

Denes		Accuracy		Input Impedance	Overload Protection
Range	50/60Hz	40 to 500Hz	500Hz to 1kHz	Input Impedance	Overload Protection
600.0mV				10MΩ<200pF	
6.000V			1.5+5	11MΩ<50pF	1000V rms AC
60.00V	0.5+5	1+5	1.5+5		1000V Pms AC
600.0V				10MΩ<50pF	10000 DC
1000V			-		

 Accuracy: At 5 to 100% of range and 1000V range is 200 to 1000V. less than 1500V peak For non-sinusoidal waveforms, add ±(2% + 2% of full scale), for Crest factor<3.</td>

 CMRR: 60dB or more DC to 60Hz (Rs=1kΩ)
 4 counts or less is corrected to 0, Response time: 2 sec max.

#### Resistance Measurement(Ω)

Range	Accuracy	Maximum Measuring Current	Open Circuit Voltage	Overload Protection
600.0Ω		<1.2mA	<3.5V	
6.000kΩ	0.4+1	<110µA		1
60.00kΩ	0.4+1	<13µA	1	
600.0kΩ		<1.3µA	<1.3V	1000V rms
6.000MΩ	0.5+1		1 <1.3V	
60.00MΩ	1+2(0 to 40MΩ)	<130nA		
00.00MΩ	2+2(40 to 60MΩ)			

– Accuracy is specified after zero adjustment at 600Ω to 6kΩ (Resistance) Response time: 2 sec max. at 600Ω to 600kΩ, 10 sec max. at 6M to 60MΩ

#### Continuity Check(刎)

Range	Range of Operation	Measuring Current	Open Circuit Voltage	Overload Protection
600.0Ω	Buzzer sounds at lower than 50±30Ω	Approx. <1.2mA	<3.5V	1000V rms

DC Current Measurement(...)(A)

Range	Accuracy	Voltage Drop	Overload Protection
600.0µA		<0.12mV/µA	
6000µA	0.2+2	<0.12mv/µA	440mA Protected by a
60.00mA		<3.3mV/mA	440mA/1000V fuse.
440.0mA		<3.3mv/mA	
6.000A	0.5+5	<0.1V/A	10A Protected by A
10.00A		<0.1V/A	10A/1000V fuse.

#### AC Current Measurement[True RMS]( $\sim$ A)

AC Current Measure	ment[True RMS]( $\sim$	A)	True RMS	value detection, sine wave		
D	Accuracy		V IL D	Overland Protection		
Range	50/60Hz	40Hz to 1kHz	Voltage Drop	Overload Protection		
600.0µA			10.10.14/			
6000µA	1	+5 1.5+5	<0.12mV/µA	440mA Protected by a		
60.00mA	0.75+5			1.6.6	<3.3mV/mA	440mA/1000V fuse.
440.0mA	0.75+5		<3.3mV/mA			
6.000A				-0.11//4	10A Protected by A	
10.00A			<0.1V/A	10A/1000V fuse.		

Accuracy: At 5 to 100% of range, 10A range is 2 to 10A and 440mA range is 30 to 440mA. For non-sinusoidal waveforms, add ±(2% + 2% of full scale), for Crest factor<3. 4 counts or less is corrected to 0, Response time: 3 sec max.

#### Diode Test(+X+)

Range	Accuracy	Measuring Current(Vf=0.6V)	Open Circuit Voltage	Overload Protection
2.000V	1+2	Approx. 0.5mA	<3.5V	1000V rms

#### Temperature Measurement(TEMP)

Range	Accuracy	Overload Protection				
-50.0 to 600.0°C	2+2°C	1000V rms				
Use optional Temperature Probe: Thermocouple Type K						

Range	Accuracy	Input Voltage
10.00 to 99.99Hz		0.2 to 600V rms
90.0 to 999.9Hz	0.00.1	0.2 to 6000 rms
0.900 to 9.999kHz	0.02+1	0.4 to 600V rms
9.00 to 99.99kHz		0.8 to 100V rms

Capacitor Measurement(++)

Range	Accuracy	Overload Protection
10.00nF	2+10	
100.0nF		
1.000µF	2+5	1000V rms
10.00µF	]	1000 mms
100.0µF	3+5	
1000µF	] 3+3	

Accuracy is specified after zero adjustment at 10nF to  $1\mu F$  (Capacitance).

#### Selection Guide

Model	1051	1052	1061	1062
Display		-		-
Detection method	True RMS	True RMS/MEAN	True RMS	True RMS/MEAN
Maximum count display	6000	6000	50000	50000
Dual display	•	•	•	•
Bar graph	31-segment	31-segment	51-segment	51-segment
Back light	White LED	White LED	White LED	White LED
Function				
Auto hold	•	•	•	•
Peak hold	-	-	-	•
Max/Min/Avg	-	•	•	•
REL	•	•	•	•
Manual memory	-	•	•	•
Logging memory	-	•	•	•
Communication	-	•	•	•
Frequency response	40Hz to 1kHz	40Hz to 1kHz	10Hz to 20kHz	10Hz to 100kHz
Operating temperature	−10 to 55℃	−10 to 55℃	−20 to 55℃	−20 to 55°C
Safety standard	CAT Ⅲ 1000V CAT Ⅳ 600V	CAT II 1000V CAT IV 600V	CAT Ⅲ 1000V CAT Ⅳ 600V	CAT Ⅲ 1000V CAT Ⅳ 600V

Model	1051	1052	1061	1062
Measurement				
DC Voltage	600.0mV to 1000V	600.0mV to 1000V	50.000mV to 1000.0V	50.000mV to 1000.0V
AC Voltage	600.0mV to 1000V	600.0mV to 1000V	500.00mV to 1000.0V	50.000mV to 1000.0V
DC Current	600.0µA to 10.00A	600.0µA to 10.00A	500.00µA to 10.000A	500.00µA to 10.000A
AC Current	600.0µA to 10.00A	600.0µA to 10.00A	500.00µA to 10.000A	500.00µA to 10.000A
AC + DC	-	-	•	•
Resistance	600.0Ω to 60.00MΩ	600.0Ω to 60.00MΩ	500.00Ω to 50.000MΩ	500.00Ω to 50.000MΩ
Frequency	10.00Hz to 99.99kHz	10.00Hz to 99.99kHz	2.000Hz to 99.99kHz	2.000Hz to 99.99kHz
Temperature	−50.0 to 600.0℃	−50.0 to 600.0℃	−200.0 to 1372.0℃	-200.0 to 1372.0°C
Capacitance	10.00nF to 1000µF	10.00nF to 1000µF	5.000nF to 50mF	5.000nF to 50mF
Duty cycle	-	-	•	•
Decibel calculation	-	-	•	•
Continuity Check	•	•	•	•
Diode Test	•	•	•	•
Low power-Ω	-	-	-	•

#### Top Class Digital Multimeters KEW 1061/1062

#### General Specification

Measurement function:	DC Voltage, AC Voltage, DC Current, AC Current, Resistance, Frequency, Temperature, Capacitor, Duty cycle ratio, Decibel (dBv, dBm), Continuity Check,	Measurement cycle:	6 times per second (except frequency measurement: one time per second, Resistance measurement : four times per second, capacitor measurement (50mF):
	Diode Test Low power- $\Omega^*$ , Effective value (root mean square value) detection		max. 0.03 time per second) Bar graph display 15 times per second
	(True RMS) and mean value detection (MEAN) can be switched during AC volt-	Operating temperature	
		Operating temperature of	
	age or AC current measurement (1062 only).	c	-20 to 55°C, 80%RH or less (no condensation), 70%RH or less at 40 to 55°C.
	The low-pass filter can be switched on/off during AC voltage or AC current mea-		d humidity range: -40 to 70°C, 70%RH or less (no condensation)
	surement (1062 only).	lemperature coefficient:	(Accuracy at 23±5℃×0.05)/℃ or less
Other functions:	Data Hold (D+H), Auto Hold (A+H), Peak Hold* (P+H), Range Hold (R+H), Maximum		(Temperature range: −20 to 18℃ and 28 to 55℃)
	value (MAX), Minimum value (MIN), Average value (AVG), Zero Adjustment (Capacitor,	Power supply:	R6/LR6(1.5V): 4
	Resistance), Relative values, Save to Memory, Auto Power Off (Approx. 20 minutes), LCD	Battery life:	Approximately 100 hours
	backlight. *: For model KEW1062 only		(Operating hours of alkaline batteries when in DC voltage-mode.)
Display:	5-digit (LCD)7-segment		Note: The battery life varies depending on the operating conditions.
	Main-display50000 counts	Withstand voltage:	6.88kV rms AC for five seconds (across input terminals and casing)
	Sub-display50000 counts	Dimension:	Approx. 192(L)×90(W)×49(D)mm
	Bar graph indicator51-segment	Weight:	Approx. 570g (including batteries and blank cover)
	Polarity Indicator:"—" Appears automatically when the polarity is negative.	Applicable standards:	IEC 61010-1 CAT IV 600V / CAT III 1000V Pollution degree 2, IEC 61010-031,
	Overrange Indicator " OL "		IEC 61326-1(EMC)
	Low-battery Indicator" 🚛 " Appears when the batteries become low.	Accessories:	Batteries : 4, Test leads: 1set (7220A), Fuse (included): 440mA/1000V (8926), 10A/1000V (8927), Blank cover, Instruction manual: 1

#### Specification

Test conditions: Temperature and humidity: 23±5°C at 80%RH or less Accuracy: ± (% of reading + digits) Not

#### DC Voltage Measurement(....V)

Range	Accuracy 1061,1062	Input Impedance	Overload Protection	
50.000mV	0.05+10			
500.00mV	0.02+2	Approx. 100MΩ		
2400.0mV	0.02+2		1000V DC 1000V rms AC	
5.0000V	0.025+5			
50.000V		10MO	1000v rms AC	
500.00V	0.03+2	10/M12		
1000.0V				

NMRR: 80dB or more 50/60Hz ±0.1% (70dB or more 50/60Hz ±0.1% when 50mV Range) CMRR: 100dB or more 50/60Hz (Rs=1k $\Omega)$   $\ Response time: 0.3 sec. max.$ 

#### ...

Denes		Upper:10	61; Lower:1	062; -:No	Specified		Input	Overload
Range	10 to 20Hz	20Hz to 1kHz	1k to 10kHz	10k to 20kHz	20k to 50kHz	50k to 100kHz	Impedance	Protection
50.000mV	-	-	-	-		_		
50.000mv	2+80 <sup>®2</sup>	0.4+40 <sup>®2</sup>	5+40 <sup>®2</sup>	5.5+40 <sup>®2</sup>	15+	40 <sup>®2</sup>	11MΩ<50pF	
500.00mV							TIM02<30pr	1000V rms
5.0000V	1.5+30 <sup>®1</sup>	0.7+	30 <sup>®1</sup>	2+50 <sup>®2</sup>	-	-		
50.000V	1+30 <sup>±1</sup>	0.4+	30 <sup>®1</sup>	1+40 <sup>±1</sup>	2+70 <sup>®2</sup>	5+200 <sup>®2</sup>		AC 1000V DC
500.00V	1						10110-50 5	10000000
1000.01/	*2	<b>※</b> 2	3+30 <sup>®2</sup>		-		10MΩ<50pF	
1000.0V	*2	*2	3+30 <sup>®2</sup>		-			

%1: At 5 to 100% of range %2: At 10 to 100% of range Crest factor <1.5V at 1000V range; Crest factor <3 at other range

CMRR: 80dB or more DC to 60Hz (Rs=1kΩ)Response time: 1 sec max

#### AC Voltage Measurement [MEAN] (~V) %1062 only AC Coupling, True RMS value detection, sine wave

Denne		Accuracy		In suit In suid an an	Overload Protection	
Range	10 to 20Hz	20 to 500Hz	500 to 1kHz	Input Impedance	Overload Protection	
50.000mV	4+80 <sup>#2</sup>	1.5+30 <sup>#2</sup>	5+30 <sup>®2</sup>			
500.00mV				11MΩ<50pF	1000V rms AC	
5.0000V	2+30**1	1+30**1	3+30*1			
50.000V	2+30	1+30	3+30		1000V DC	
500.00V				10MΩ<50pF		
1000.0V	*2	*2	<b>※</b> 2	]		

※1: At 5 to 100% of range ※2: At 10 to 100% of range

CMRR: 80dB or more DC to 60Hz (Rs=1kΩ) Response time: 1 sec max

#### DCV+ACV(<del>…</del> + ∼)

DCV+ACV(+~) AC Coupling, True RMS							value detectio	n, sine wave
	Accuracy (Upper:1061; Lower:1062; -:Not Specified)			Overload				
Range	DC,10 to 20Hz	DC,20Hz to 1kHz	DC,1k to 10kHz	DC,10k to 20kHz	DC,20k to 50kHz	DC,50k to 100kHz	Input Impedance	Protection
5.0000V	1.5+10**1	1.1	o <sup>楽1</sup>	2+10 <sup>⊕2</sup>			11MΩ<50pF	
50.000V	1.5+10 1.5+10 <sup>®1</sup>		1+10 <sup>®1</sup> 0.5+10 <sup>®1</sup>	2+10 <sup>®2</sup>	5+20 <sup>®2</sup>		1000V rms	
500.00V	1.3+10	0.54		1+10	2+10	10 5+20	10MΩ<50pF	AC
1000.0V	<b>※</b> 2	<b>※</b> 2		-			TOM12<30pr	1000V DC
1000.00	*2	*2		-	-			

%1: At 5 to 100% of range

X: At 10 to 100% of range Crest factor <1.5V at 1000V range; Crest factor <3 at other range CMRR: 80dB or more DC to 60Hz (Rs=1kΩ) Response time: 2 sec max.

#### Resistance Measurement(Ω)

Range	Accuracy		Maximum	Open Circuit	Overload		
Kulige	1061	1062	Measuring Current	Voltage	Protection		
500.00Ω		0.05+2**1	<1mA				
5.0000kΩ	0.1+2**		0.05+2*1 <0.25mA <25µA				
50.000kΩ	0.1+2			0.0312	0.0312	0.0312	<25µA
500.00kΩ	1		<2.5µA	<2.5V	1000V rms		
5.0000MΩ	0.5	5+2	<1.5µA	1			
50.000MΩ	1.	+2	<0.13µA	1			

-Accuracy is specified after zero adjustment (resistance). Response time: 1 sec. max. at 500Ω to 500kΩ, 5 sec. max. at 5 to 50MΩ

$LowPower-\Omega(LP-\Omega)$	%1062 only	Max	imum Reading 5000		
Range	Accuracy	Maximum Measuring Current	Open Circuit Voltage	Overload Protection	
5.000kΩ		<10µA			
50.00kΩ	0.2+3	<1.0µA	<0.7V	1000V rms	
500.0kΩ		<0.6µA	<0.7 V	TOODV rms	
5.000MΩ	1+3	<0.05µA			

Continuity Check()) Maximum Reading 500							
Range	Range of Operation 1061,1062	Measuring Current	Open Circuit Voltage	Overload Protection			
500.0Ω	Buzzer sounds at lower than 100±50Ω	Approx. 0.5mA	<5V	1000V rms			

te: Each response time is a value to rated accuracy within selected range.								
	DC Current Measureme	nt( <u></u> )(A)						
	Range	Accuracy 1061,1062	Voltage Drop	Overload Protection				
	500.00µA		<0.11mV/µA					
	5000.0µA	0.2+5	<0.11mv/μA	440mA Protected by a				
	50.000mA	7 0.2+3	<4mV/mA	440mA/1000V fuse.				
	500.00m A	ר I	<4mv/mA					

500.00µA		<0.11mV/µA		
5000.0µA	0.2+5	<0:11mt/μA	440mA Protected by a	
50.000mA	0.215	<4mV/mA	440mA/1000V fuse.	
500.00mA		<4mv/mA		
5.0000A	0.6+10	<0.1V/A	10A Protected by A	
10.000A	0.6+5	\$0.1V/A	10A/1000V fuse.	
Maximum measurement cur	rent : 440mA at 500mA range	1		

Response time: 0.3 sec. max.

C Current Measurement [True RMS](~A)			True	RMS value dete	ction, sine wave
Range	Upper:1061; Lower:1062; -:Not Specified			Voltage Drop	Overload
Kulige	10 to 20Hz	20Hz to 1kHz	1k to 5kHz	1 volidge brop	Protection
500.00µA				<0.11mV/µA	
5000.0µA	1.5+20	1+20	-	<0.11mv/μA	440mA Protected by a
50.000mA	1+20	0.75+20	1+30	<4mV/mA	440mA/1000V fuse.
500.00mA	1			<4mv/mA	
5.0000A	1.5+20	1+20	-	<0.1V/A	10A Protected by A
10.000A	1.5+20	1+20	2+30	<0.1V/A	10A/1000V fuse.

440mA at 500mA range Crest factor<3. Response time: 1 sec max.

#### AC Current Measurement [MEAN] (~A) %1062 only MEAN value detection, True RMS value calibration

Range	Accuracy			Voltage Drop	Overload Protec-
Kange	10 to 20Hz	20 to 500Hz	500Hz to 1kHz	voliage Drop	tion
500.00µA				<0.11mV/µA	440mA Protected by a 440mA/1000V fuse.
5000.0µA	2+20	1 5 . 20	+20 2+30	<0.11mm/μA	
50.000mA	2+20	1.5+20		<4mV/mA	
500.00mA					<4mv/mA
5.0000A	3+20	0.00	4+30	<0.1V/A	10A Protected by A
10.000A	3+20	2+20	4+30	<0.1V/A	10A/1000V fuse.

Accuracy At 5 to 100% of range, At 10 to 100% of range for 10A Range 440mA at 500mA range Response time: 1 sec max.

#### DCA+ACA(...+~)

DCA+ACA(+~) Maximum Reading 5000								
Deres	Accuracy (Uppe	r:1061; Lower:1062;	-:Not Specified)	Voltage Drop	Overload			
Range	DC,10 to 20Hz	DC,20Hz to 1kHz	DC,1k to 5kHz	voliage Drop	Protection			
500.00µA				<0.11mV/µA				
5000.0µA	2+10	1.5+10	_	<0.11mv/μA	440mA Protected by a			
50.000mA	1.5+10	1+10	1.5+10	<4mV/mA	440mA/1000V fuse.			
500.00mA	1			<4mv/mA				
5.0000A	2+10	1.5+10	-	<0.1V/A	10A Protected by A			
10.000A	2+10	1.5+10	3+10	<0.1V/A	10A/1000V fuse.			

Accuracy At 5 to 100% of range, At 10 to 100% of range for 10A Range

440mA at 500mA range Crest factor<3. Response time: 2 sec max.

#### Diode Test(-K-)

Range	Accuracy 1061,1062	Measuring Current (Vf=0.6V)	Open Circuit Voltage	Overload Protection
2.4000V	1+2	Approx. 0.5mA	<5V	1000V rms

Temperature Measurement(TEMP) 
 Range
 Accuracy 1061,1062
 Over

 -200.0 to 1372.0°C
 1+1.5°C
 10
 ad Protection 1000V rms

#### Use optional Temperature Probe: Thermocouple Type K

Capacitor Measurement(HF) Maximum Read							
	Range	Accuracy 1061,1062	Overload Protection				
	5.000nF						
	50.00nF						
	500.0nF	1+5*1					
	5.000µF	1	1000V rms				
	50.00µF		1000v rms				
	500.0uF	2+5					

5.000m 3+5 %1: Accuracy is specified after zero adjustment (capacitor)

Frequency Measurement(Hz) AC Coupling, Maximum Reading 9999

Kange (AUTO)	
2.000 to 9.999Hz	
9.00 to 99.99Hz	0.02+1*1
90.0 to 999.9Hz	0.02+1
0.900 to 9.999kHz	
9.00 to 99.99kHz	*2

At 10 to 100% of input voltage or current range \*2: At 40 to 100% of input voltage or current range

#### Duty cycle ratio(%)

Range 10 to 90% Accuracy 1061,1062 % 1: At 10.00 to 500.0Hz, square wave At 40 to

100% of input voltage or current range

Peak Hold(P∙H)							
Range	Range Resolution Response Time Maximum						
DCV, DCA ±100 digit >250µs							

#### Accessories

Description	MODEL	Contents
Test leads	7220A	CAT IV 600V / CAT III 1000V 1set
Free	8926	440mA/1000V×1
Fuse	8927	10A/1000V×1



#### Optional accessories

Description	MODEL	Contents
Alligator clip	7234	CAT IV 600V / CAT III 1000V 1set
USB Communication set	8241	USB adapter+USB cable
	8405	-40 to 500°C (Surface type, Point material: Ceramic)
The second states of K	8406	-40 to 500°C (Surface type)
Thermocouple Type K	8407	-40 to 700°C (Liquid, Semi-solid)
	8408	-40 to 600°C (Air, Gas)
	8115	130A AC / 180A DC
	8121	100A AC
	8122	500A AC
Clamp sensor	8123	1000A AC
	8146	30A AC
	8147	70A AC
	8148	100A AC
Banana Ø4mm adjuster plug	7146	Length :190mm
Carrying case	9154	Soft case (for the main unit with test leads and communication cable)

# 7234 Alligator clip 7234 Second particular plug 105B Comminication set USB Comminication set 0154 Second particular plug 9154 Second plug Carrying case Ti46 Banana Ø4 adjuster plug

#### Clamp sensor Specification

	AC/DC current sensor		AC current sensor			Leakage & AC current sensor		
MODEL	8115	8121	8122	8123	8146	8147	8148	
	CE	<b>P</b>	Ce Ce	<b>V</b>	<b>P</b>			
Conductor size	φ12	φ24	<i>φ</i> 40	φ55	φ24	φ40	φ68	
Rated current	130A AC / 180A DC	100A AC	500A AC	1000A AC	30A AC	70A AC	100A AC	
Output voltage	10mV/A AC/DC	500mV/100A AC	500mV/500A AC	500mV/1000A AC	1500mV/30A AC	3500mV/70A AC	5000mV/100A AC	
Accuracy (50/60Hz)	AC ±1.2%rdg±0.4mV DC ±1.2%rdg±0.4mV (This accuracy is defined after a zero-adjustment)		±2.0%rdg±0.3mV		0 to 15A ±1.0%rdg±0.1mV 15 to 30A ±5.0%rdg	0 to 40A ±1.0%rdg±0.1mV 40 to 70A ±5.0%rdg	0 to 80A ±1.0%rdg±0.1mV 80 to 100A ±5.0%rdg	
Frequency range				40Hz to 1kHz				
Dimension	127(L)×42(W)×22(D)mm	97(L)×59(W)×26(D)mm	128(L)×81(W)×36(D)mm	170(L)×105(W)×48(D)mm	100(L)×60(W)×26(D)mm	128(L)×81(W)×36(D)mm	186(L)×129(W)×53(D)mm	
Weight	Approx. 140g	Approx. 150g	Approx. 260g	Approx. 360g	Approx. 150g	Approx. 240g	Approx. 510g	

#### • Thermocouple Type K Specification

Model	Usage	Measurement temprature	Tolerance (t: measurement temperature)	Response speed
8405	Surface type (Point material: Ceramic)	-40 to 500°C	±2.5°C/t=-40 to 333°C,	Approx. 1.8 Sec.
8406	Surface type	-40 10 5000	±0.0075×  +  °C/ t=333 to 500°C	Approx. 1.0 Sec.
8407	Liquid, Semi-solid	-40 to 700℃	±2.5℃/t=-40 to 333℃, ±0.0075× t ℃/t=333 to 700℃	1 Sec. or less
8408	Air, Gas	-40 to 600°C	±2.5℃/t=-40 to 333℃, ±0.0075× t ℃/ t =333 to 600℃	0.4 Sec.





Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely for correct use. Failure to follow the safety rules can cause fire, trouble, electrical shock, etc. Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

#### For inquiries or orders :



2-5-20, Nakane, Meguro-ku, Tokyo, 152-0031 Japan Phone:+81-3-3723-0131 Fax:+81-3-3723-0152

www.kew-ltd.co.jp